

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130299</u>
Name <u>Chris Babcock</u>		ID# <u>3867</u> Date <u>12-8-09</u>
A	Agency <u>King Cove DPS</u>	Phone # <u>497-2210</u>
Instrument Location <u>Squad Room</u>		
Alco S/N <u>X301600</u> Target Value <u>0.080</u> High Pressure <u>900</u>		
B	Alco Test Value Average <u>.080</u> <u>.080</u> 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u>		
(OVER)		


AMB
12/21/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of January, 2010
[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299 ✓

DECEMBER 08, 2009

OPERATOR'S NAME:
BARCOCK CHRISTOPHER M
OPERATOR'S NUMBER: 3867
SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 123456
DEPT/AGENCY: KUC1
CASE/REPORT: UOC
TEST TYPE: U

ALCO TARGET VALUE: .000 ✓
ALCO S/N: K301633

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.27 in
ALCO TARGET .000 13:23
BLANK TEST .000 13:24
INTERNAL STANDARD VERIFIED 13:24
ALCO TU 30.29 in .000 ✓ 13:24
BLANK TEST .000 13:25
SUBJECT SAMPLE .000 13:26
BLANK TEST .000 13:26
ALCO TU 30.27 in .000 ✓ 13:27
BLANK TEST .000 13:28

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299 ✓

DECEMBER 08, 2009
TIME 13:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.29 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

"##%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\\]^_`{|}~!@#\$%^&'()*+
,-./0123456789:;<=>?@ABCDEFGHIJ